


## CT-6 Leg Splint Visual User Guide



FareTec Inc  
1610 West Jackson St  
Painesville OH 44077  
440 350 9510 Voice  
440 350 9520 Fax  
800 322 3273 Toll Free

Copyright ©2010 Fare Tec. Inc.

### Assess Injury

### Assemble


#### Step 1

Check circulation

Cut off pant leg if needed.



Assess injury to determine if traction is needed.  
Determine pain level. Compare lengths of the patient's legs.



#### Step 2

Remove splint from bag.



### Components

### Assemble

### Sizing




#### Step 3

Shake splint up and down to connect the tube sections.  
Manually connect any unlinked sections.



#### Step 4

Place unit alongside uninjured leg, extend just above hip crest and patient's foot.



Two straps above knee. Two straps above hip crest.

#### Repacking:


1. Slide leg straps to end of splint, but not over tube edges.
2. Pull apart tube sections and fold each on top of the other. Allow straps to hang.
3. Fold ankle hitch and ischial strap alongside folded tube sections.
4. Roll up splint tightly with leg straps and slide into bag, ankle hitch first.

#### Care and Use:


- USE PPE** : Gloves, mask and eye protection.  
**CLEAN** : Wash with soap and water first and dry with disposable towels.  
**DISINFECT** : Use commercial disinfectant according to directions or solution of 1 part bleach to 100 parts water. Treat entire CT-6 and dry with disposable towels.  
**DISCARD** : Dispose of used cleaning materials and all PPE (except eye protection) in properly

## Adjust Length

**Step 5** Add or remove tube sections as necessary for proper sizing. Fold any disconnected ones alongside as shown.




Secure disconnected sections with ischial cap.




## Attach Strap

**Step 6** Wrap ischial strap around leg and attach clip.



Ensure buckle is on top, not under leg. Tighten buckle strap.




## Attach Ankle Hitter

**Step 7** Place foot into ankle




Use sliding end of foot strap to ensure positioned on both sides of ankle. V around ankle and fold small tab. Tighten foot strap buckle to eliminate




## Traction

**Step 8** Pull line to apply moderate tension. This can be achieved by pulling the line exiting the purchase block. Minimal amount of traction should be applied so that the splint is resting in its appropriate position.




Lift the line up into the V-jam to hold tension.




## Securing and Applying Traction

**Step 9** Wrap leg straps, two above knee and two below. Do not cover injury or knee.




Apply traction by re-adjusting the tension until patient's comfort is achieved. Lift the line up and relock into the V-jam.



## Final Steps

**Step 10** Tuck excess line



### Finished view of CT-6



### Basic Guidelines For Applying Tra

- 1) Apply traction until the patient's pain
- 2) Apply traction until the patient's legs

Training video available at [www.faretec.com](http://www.faretec.com)